

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072120\
 Data File : VN062576.D
 Acq On : 21 Jul 2020 17:15
 Operator : JC/MD
 Sample : L3342-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Jul 22 05:39:49 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 21 18:48:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	180319	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	315365	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	301315	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	107575	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	140511	50.16	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.32%	
35) Dibromofluoromethane	7.55	113	104898	50.64	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.28%	
50) Toluene-d8	10.06	98	408869	49.47	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.94%	
62) 4-Bromofluorobenzene	12.38	95	139385	46.12	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.24%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN072120\
Data File : VN062576.D
Acq On : 21 Jul 2020 17:15
Operator : JC/MD
Sample : L3342-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Jul 22 05:39:49 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 21 18:48:59 2020
Response via : Initial Calibration

