


2-Hexanone

Response Ratio




Resp Ratio = 4.34×10^{-1} * Amt - 2.57×10^{-2}
Coef of Det (r^2) = 1.000 Curve Fit: Linear

Method Name: Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
Calibration Table Last Updated: Tue Jul 21 14:03:23 2020


1,2,4-Trichlorobenzene

Response Ratio



Resp Ratio = $8.94 \times 10^{-1} \times \text{Amt} - 5.55 \times 10^{-2}$
Coef of Det (r^2) = 0.997 Curve Fit: Linear


Method Name: Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
Calibration Table Last Updated: Tue Jul 21 14:03:23 2020



Method Name: Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
Calibration Table Last Updated: Tue Jul 21 14:03:23 2020

Naphthalene

Response Ratio




Resp Ratio = 2.40×10^0 * Amt - 1.81×10^{-1}
Coef of Det (r^2) = 0.996 Curve Fit: Linear

Method Name: Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
Calibration Table Last Updated: Tue Jul 21 14:03:23 2020


Acrolein

Response Ratio



Resp Ratio = $1.33\text{e-}001 \times \text{Amt} - 2.25\text{e-}002$
Coef of Det (r^2) = 1.000 Curve Fit: Linear


Method Name: Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
Calibration Table Last Updated: Tue Jul 21 14:03:23 2020



Method Name: Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
Calibration Table Last Updated: Tue Jul 21 14:03:23 2020

2-Chloroethyl Vinyl ether

Response Ratio



Resp Ratio = $2.11 \times 10^{-1} \times \text{Amt} - 7.03 \times 10^{-2}$
Coef of Det (r^2) = 0.998 Curve Fit: Linear

Method Name: Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
Calibration Table Last Updated: Tue Jul 21 14:03:23 2020