

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072121\
 Data File : VN067837.D
 Acq On : 21 Jul 2021 14:36
 Operator : JC/MD
 Sample : M3038-01
 Misc : 5.07g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-06-5.5-6.0

Manual Integrations
 APPROVED

MMDadoda
 7/22/2021 6:03:16 PM

Quant Time: Jul 21 15:11:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 07 16:18:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	331563	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	620622	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	564567	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	235015	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	258198	51.527	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	103.060%
35) Dibromofluoromethane	8.016	113	181881	48.506	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.020%
50) Toluene-d8	10.443	98	775657	49.420	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.840%
62) 4-Bromofluorobenzene	12.736	95	256376	46.241	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	92.480%
Target Compounds						Qvalue
18) Methyl Acetate	4.865	43	21584m	4.874	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072121\
Data File : VN067837.D
Acq On : 21 Jul 2021 14:36
Operator : JC/MD
Sample : M3038-01
Misc : 5.07g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SB-06-5.5-6.0

Manual Integrations
APPROVED

MMDadoda
7/22/2021 6:03:16 PM

Quant Time: Jul 21 15:11:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 07 16:18:53 2021
Response via : Initial Calibration

