

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072121\
 Data File : VN067839.D
 Acq On : 21 Jul 2021 15:25
 Operator : JC/MD
 Sample : PB137834TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB137834TB

Manual Integrations
 APPROVED

MMDadoda
 7/22/2021 6:03:20 PM

Quant Time: Jul 22 02:12:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 07 16:18:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	326504	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	614517	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	644093	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	231758	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	255151	51.708	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 103.420%			
35) Dibromofluoromethane	8.024	113	186371	50.197	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 100.400%			
50) Toluene-d8	10.443	98	804031	51.737	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 103.480%			
62) 4-Bromofluorobenzene	12.734	95	273563	49.831	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 99.660%			
Target Compounds					Qvalue	
16) Acetone	4.296	43	51468	33.148	ug/l #	88
18) Methyl Acetate	4.865	43	6944m	1.592	ug/l	
20) Methylene Chloride	5.098	84	28220	7.190	ug/l	88
43) Isopropyl Acetate	8.566	43	44236	4.547	ug/l #	92
95) Naphthalene	15.509	128	16988	4.419	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072121\
Data File : VN067839.D
Acq On : 21 Jul 2021 15:25
Operator : JC/MD
Sample : PB137834TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB137834TB

Manual Integrations
APPROVED

MMDadoda
7/22/2021 6:03:20 PM

Quant Time: Jul 22 02:12:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 07 16:18:53 2021
Response via : Initial Calibration

