

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072121\
 Data File : VN067847.D
 Acq On : 21 Jul 2021 18:41
 Operator : JC/MD
 Sample : M3089-19
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 R99997-C-TOP-(0)

Quant Time: Jul 22 02:14:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 07 16:18:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	295009	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	572049	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	556443	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	198477	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	237033	53.164	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 106.320%			
35) Dibromofluoromethane	8.024	113	171347	49.577	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 99.160%			
50) Toluene-d8	10.446	98	726897	50.246	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 100.500%			
62) 4-Bromofluorobenzene	12.733	95	242434	47.439	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 94.880%			
Target Compounds					Qvalue	
16) Acetone	4.296	43	107590	76.691	ug/l	93
18) Methyl Acetate	4.864	43	11923	3.026	ug/l #	84
20) Methylene Chloride	5.090	84	33224	9.582	ug/l	91
43) Isopropyl Acetate	8.566	43	25664	2.834	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072121\
Data File : VN067847.D
Acq On : 21 Jul 2021 18:41
Operator : JC/MD
Sample : M3089-19
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
R99997-C-TOP-(0)

Quant Time: Jul 22 02:14:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 07 16:18:53 2021
Response via : Initial Calibration

