

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072121\
 Data File : VN067848.D
 Acq On : 21 Jul 2021 19:06
 Operator : JC/MD
 Sample : M3089-20
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 R99997-C-BOTTOM-(3)

Manual Integrations
 APPROVED

MMDadoda
 7/22/2021 6:03:24 PM

Quant Time: Jul 22 02:14:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 07 16:18:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	305807	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	595460	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	602928	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	225585	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	246405	53.315	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 106.640%			
35) Dibromofluoromethane	8.024	113	179560	49.911	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 99.820%			
50) Toluene-d8	10.443	98	772021	51.267	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 102.540%			
62) 4-Bromofluorobenzene	12.734	95	267312	50.250	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 100.500%			
Target Compounds					Qvalue	
16) Acetone	4.296	43	43278	29.760	ug/l	97
18) Methyl Acetate	4.867	43	10326m	2.528	ug/l	
20) Methylene Chloride	5.095	84	33667m	9.351	ug/l	
43) Isopropyl Acetate	8.568	43	35420	3.757	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072121\
Data File : VN067848.D
Acq On : 21 Jul 2021 19:06
Operator : JC/MD
Sample : M3089-20
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
R99997-C-BOTTOM-(3)

Manual Integrations
APPROVED

MMDadoda
7/22/2021 6:03:24 PM

Quant Time: Jul 22 02:14:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 07 16:18:53 2021
Response via : Initial Calibration

