

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072122\
 Data File : VN073548.D
 Acq On : 21 Jul 2022 11:45
 Operator : JC\MD
 Sample : VN0721MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0721MBL01

Quant Time: Jul 21 15:48:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	249414	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	394300	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	359470	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	161617	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	151166	41.984	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	83.960%
35) Dibromofluoromethane	7.951	113	132141	50.371	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.740%
50) Toluene-d8	10.380	98	415006	44.542	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.080%
62) 4-Bromofluorobenzene	12.669	95	166055	46.377	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072122\
Data File : VN073548.D
Acq On : 21 Jul 2022 11:45
Operator : JC\MD
Sample : VN0721MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0721MBL01

Quant Time: Jul 21 15:48:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

