

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072123\
 Data File : VN078609.D
 Acq On : 21 Jul 2023 13:22
 Operator : JC\MD
 Sample : 03709-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-01-072023

Quant Time: Jul 21 23:21:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 10 16:13:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	423529	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	742047	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	678508	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	243710	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	272714	48.031	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.060%
35) Dibromofluoromethane	8.177	113	269617	53.525	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.040%
50) Toluene-d8	10.571	98	879229	46.204	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.400%
62) 4-Bromofluorobenzene	12.853	95	301625	51.460	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.920%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072123\
Data File : VN078609.D
Acq On : 21 Jul 2023 13:22
Operator : JC\MD
Sample : 03709-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-01-072023

Quant Time: Jul 21 23:21:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 10 16:13:53 2023
Response via : Initial Calibration

