

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072123\
 Data File : VN078616.D
 Acq On : 21 Jul 2023 16:12
 Operator : JC\MD
 Sample : 03704-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Jul 21 23:26:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 10 16:13:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	416009	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	749159	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	693822	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	234873	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.589	65	276792	49.631	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.260%
35) Dibromofluoromethane	8.177	113	276110	54.293	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	108.580%
50) Toluene-d8	10.571	98	887733	46.208	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.420%
62) 4-Bromofluorobenzene	12.853	95	302804	51.170	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072123\
Data File : VN078616.D
Acq On : 21 Jul 2023 16:12
Operator : JC\MD
Sample : 03704-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Jul 21 23:26:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 10 16:13:53 2023
Response via : Initial Calibration

