

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072123\
 Data File : VN078617.D
 Acq On : 21 Jul 2023 16:36
 Operator : JC\MD
 Sample : 03664-01ME
 Misc : 5.04g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AUD-1021-1020-COMPME

Quant Time: Jul 21 16:54:05 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 10 16:13:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	426246	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	749674	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	736420	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	374814	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	272201	47.635	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.280%
35) Dibromofluoromethane	8.171	113	280492	55.117	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	110.240%
50) Toluene-d8	10.571	98	868616	45.182	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	90.360%
62) 4-Bromofluorobenzene	12.853	95	375030	63.332	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	126.660%#
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	5.512	59	21324m	24.200	ug/l	
16) Acetone	4.453	43	71637	40.246	ug/l	99
52) Toluene	10.635	92	60251	4.466	ug/l	99
68) m/p-Xylenes	12.076	106	13613	1.323	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072123\
Data File : VN078617.D
Acq On : 21 Jul 2023 16:36
Operator : JC\MD
Sample : 03664-01ME
Misc : 5.04g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AUD-1021-1020-COMPME

Quant Time: Jul 21 16:54:05 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 10 16:13:53 2023
Response via : Initial Calibration

