

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072219\
 Data File : VN056866.D
 Acq On : 22 Jul 2019 20:12
 Operator : JC/SP
 Sample : PB121556ZHE#06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB121556ZHE#06

Quant Time: Jul 23 03:32:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072219W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 23 01:46:34 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	263844	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	488772	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	534356	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	181936	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	154289	54.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.64%	
35) Dibromofluoromethane	7.59	113	146392	47.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.50%	
50) Toluene-d8	10.09	98	620884	52.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.78%	
62) 4-Bromofluorobenzene	12.40	95	222710	57.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.84%	
Target Compounds						
16) Acetone	3.82	43	56077	55.333	ug/l	96
18) Methyl Acetate	4.33	43	13769	2.395	ug/l #	71
20) Methylene Chloride	4.54	84	7656	2.526	ug/l #	91
43) Isopropyl Acetate	8.17	43	17174	2.748	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN072219\
Data File : VN056866.D
Acq On : 22 Jul 2019 20:12
Operator : JC/SP
Sample : PB121556ZHE#06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB121556ZHE#06

Quant Time: Jul 23 03:32:37 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072219W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 23 01:46:34 2019
Response via : Initial Calibration

