


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN072219\
 Data File : VN056849.D
 Acq On : 22 Jul 2019 9:53
 Operator : JC/SP
 Sample : BFB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072219W.M
 Title : SW846 8260
 Last Update : Tue Jul 23 01:46:34 2019



AutoFind: Scans 3304, 3305, 3306; Background Corrected with Scan 3290

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.2	7342	PASS
75	95	30	60	45.9	20824	PASS
95	95	100	100	100.0	45365	PASS
96	95	5	9	6.7	3046	PASS
173	174	0.00	2	1.2	480	PASS
174	95	50	100	89.7	40680	PASS
175	174	5	9	7.8	3179	PASS
176	174	95	101	99.3	40392	PASS
177	176	5	9	6.5	2641	PASS