

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072220\
 Data File : VN062595.D
 Acq On : 22 Jul 2020 15:11
 Operator : JC/MD
 Sample : PB130365TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB130365TB

Quant Time: Jul 23 07:35:44 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 21 18:48:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	166689	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	292794	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	280835	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	97013	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	133033	51.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.76%	
35) Dibromofluoromethane	7.55	113	98354	51.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.28%	
50) Toluene-d8	10.06	98	379018	49.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.78%	
62) 4-Bromofluorobenzene	12.38	95	125936	44.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.76%	
Target Compounds						
16) Acetone	3.78	43	5634	5.782	ug/l	90
20) Methylene Chloride	4.50	84	16068	6.875	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN072220\
Data File : VN062595.D
Acq On : 22 Jul 2020 15:11
Operator : JC/MD
Sample : PB130365TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB130365TB

Quant Time: Jul 23 07:35:44 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 21 18:48:59 2020
Response via : Initial Calibration

