

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072220\
 Data File : VN062596.D
 Acq On : 22 Jul 2020 15:35
 Operator : JC/MD
 Sample : PB130362TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB130362TB

Quant Time: Jul 23 07:37:25 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 21 18:48:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	169342	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	298174	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	301029	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	105816	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	137052	52.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.20%	
35) Dibromofluoromethane	7.55	113	99563	50.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.66%	
50) Toluene-d8	10.06	98	397620	50.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.76%	
62) 4-Bromofluorobenzene	12.38	95	139262	48.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.48%	
Target Compounds						
16) Acetone	3.79	43	17744	17.924	ug/l	97
18) Methyl Acetate	4.29	43	3136	1.251	ug/l #	86
20) Methylene Chloride	4.51	84	40643	17.118	ug/l	97
25) 2-Butanone	6.79	43	10079	6.685	ug/l	94
43) Isopropyl Acetate	8.13	43	66952	12.123	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN072220\
Data File : VN062596.D
Acq On : 22 Jul 2020 15:35
Operator : JC/MD
Sample : PB130362TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB130362TB

Quant Time: Jul 23 07:37:25 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 21 18:48:59 2020
Response via : Initial Calibration

