

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072221\
 Data File : VN067867.D
 Acq On : 22 Jul 2021 17:27
 Operator : JC/MD
 Sample : M3132-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 55-ST-32

Manual Integrations
 APPROVED

MMDadoda
 7/27/2021 3:39:10 PM

Quant Time: Jul 24 07:14:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N071421W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Jul 14 01:19:19 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.662	128	54957	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	8.968	114	314595	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.749	117	278548	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.437	65	141977	31.977	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	106.600%
60) 4-Bromofluorobenzene	12.736	95	115791	25.597	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	85.333%
63) Toluene-d8	10.446	98	386664	30.180	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	100.600%
Target Compounds						
4) Vinyl Chloride	2.445	62	2653	0.486	ug/l #	75
15) Acetone	4.293	58	3328m	8.391	ug/l	
22) cis-1,2-Dichloroethene	7.327	96	93283	23.782	ug/l	88
37) Trichloroethene	9.220	130	1080	0.282	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072221\
Data File : VN067867.D
Acq On : 22 Jul 2021 17:27
Operator : JC/MD
Sample : M3132-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
55-ST-32

Manual Integrations
APPROVED

MMDadoda
7/27/2021 3:39:10 PM

Quant Time: Jul 24 07:14:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N071421W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Jul 14 01:19:19 2021
Response via : Initial Calibration

