

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072221\
 Data File : VN067869.D
 Acq On : 22 Jul 2021 18:16
 Operator : JC/MD
 Sample : VSTDCCC020
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC020

Quant Time: Jul 24 07:14:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N071421W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Jul 14 01:19:19 2021
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	92	0.00
2 M	Dichlorodifluoromethane	1.533	1.584	-3.3	101	0.00
3 M	Chloromethane	1.915	1.963	-2.5	94	0.00
4 M	Vinyl Chloride	2.979	3.281	-10.1	101	0.00
5 M	Bromomethane	2.480	2.613	-5.4	94	0.00
6 M	Chloroethane	2.218	2.604	-17.4	103	0.00
7 M	Trichlorofluoromethane	5.528	6.300	-14.0	100	0.00
8 T	Diethyl Ether	1.028	1.099	-6.9	100	0.00
9	1,1,2-Trichlorotrifluoroeth	1.571	1.759	-12.0	105	0.00
10 M	1,1-Dichloroethene	1.455	1.604	-10.2	106	0.00
11	Methyl Iodide	1.924	1.905	1.0	102	0.00
12	Methyl Acetate	1.986	2.191	-10.3	108	0.00
13 M	Acrolein	0.208	0.221	-6.3	108	0.00
14 M	Acrylonitrile	0.873	0.915	-4.8	104	0.00
15 M	Acetone	0.217	0.235	-8.3	105	0.00
16 M	Carbon Disulfide	3.843	3.717	3.3	97	0.00
17	Allyl chloride	2.561	2.603	-1.6	98	0.00
18 M	Methylene Chloride	2.014	2.082	-3.4	103	0.00
19 M	trans-1,2-Dichloroethene	1.735	1.823	-5.1	104	0.00
20 T	Diisopropyl ether	6.063	6.263	-3.3	102	0.00
21 M	1,1-Dichloroethane	3.336	3.596	-7.8	104	0.00
22 M	cis-1,2-Dichloroethene	2.141	2.176	-1.6	103	0.00
23 M	tert-Butyl Alcohol	0.322	0.315	2.2	93	0.00
24 M	Methyl tert-Butyl Ether	6.009	6.213	-3.4	101	0.00
25 M	Chloroform	3.712	3.971	-7.0	104	0.00
26	Cyclohexane	2.906	2.854	1.8	97	0.00
27 s	1,2-Dichloroethane-d4	2.424	2.606	-7.5	90	0.00
28 I	1,4-Difluorobenzene	1.000	1.000	0.0	90	0.00
29	1,1-Dichloropropene	0.471	0.473	-0.4	100	0.00
30 M	2-Butanone	0.218	0.214	1.8	100	0.00
31	2,2-Dichloropropane	0.525	0.521	0.8	101	0.00
32 M	1,1,1-Trichloroethane	0.578	0.598	-3.5	105	0.00
33 M	Carbon Tetrachloride	0.494	0.489	1.0	103	0.00
34 M	Benzene	1.407	1.437	-2.1	102	0.00
35	Methacrylonitrile	0.224	0.235	-4.9	117	0.00
36 M	1,2-Dichloroethane	0.524	0.543	-3.6	98	0.00
37 M	Trichloroethene	0.366	0.372	-1.6	103	0.00
38	Methylcyclohexane	0.521	0.487	6.5	94	0.00
39 M	1,2-Dichloropropane	0.359	0.368	-2.5	103	0.00
40	Dibromomethane	0.258	0.264	-2.3	103	0.00
41 M	Bromodichloromethane	0.506	0.521	-3.0	104	0.00
42 M	Vinyl Acetate	0.867	0.887	-2.3	112	0.00
43	Ethyl Acetate	0.441	0.440	0.2	105	0.00
44	Isopropyl Acetate	0.800	0.788	1.5	100	0.00
45 T	1,4-Dioxane	0.006	0.006#	0.0	97	0.00
46	Methyl methacrylate	0.345	0.342	0.9	102	0.00
47	n-amyl Acetate	0.657	0.608	7.5	99	0.00
48 M	t-1,3-Dichloropropene	0.546	0.515	5.7	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072221\
 Data File : VN067869.D
 Acq On : 22 Jul 2021 18:16
 Operator : JC/MD
 Sample : VSTDCCC020
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC020

Quant Time: Jul 24 07:14:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N071421W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Jul 14 01:19:19 2021
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	cis-1,3-Dichloropropene	0.579	0.569	1.7	103	0.00
50 M	1,1,2-Trichloroethane	0.368	0.384	-4.3	108	0.00
51	Ethyl methacrylate	0.573	0.550	4.0	101	0.00
52	1,3-Dichloropropane	0.609	0.610	-0.2	101	0.00
53 M	Dibromochloromethane	0.385	0.366	4.9	105	0.00
54 M	1,2-Dibromoethane	0.380	0.376	1.1	103	0.00
55 M	2-Chloroethyl vinyl ether	0.111	0.094	15.3	96	0.00
56 M	Bromoform	0.270	0.226	16.3	103	0.00
57 I	Chlorobenzene-d5	1.000	1.000	0.0	87	0.00
58 M	4-Methyl-2-Pentanone	0.515	0.529	-2.7	100	0.00
59 M	2-Hexanone	0.361	0.361	0.0	99	0.00
60 S	4-Bromofluorobenzene	0.487	0.479	1.6	86	0.00
61 M	Tetrachloroethene	0.369	0.400	-8.4	107	0.00
62 M	Toluene	1.765	1.785	-1.1	98	0.00
63 S	Toluene-d8	1.380	1.413	-2.4	88	0.00
64 M	Chlorobenzene	1.057	1.064	-0.7	100	0.00
65	1,1,1,2-Tetrachloroethane	0.389	0.394	-1.3	104	0.00
66 M	Ethyl Benzene	1.953	1.937	0.8	97	0.00
67 M	m/p-Xylenes	0.745	0.736	1.2	97	0.00
68 M	o-Xylene	0.735	0.727	1.1	98	0.00
69 M	Styrene	1.216	1.200	1.3	99	0.00
70	Isopropylbenzene	1.897	1.854	2.3	96	0.00
71 M	1,1,2,2-Tetrachloroethane	0.539	0.556	-3.2	101	0.00
72	1,2,3-Trichloropropane	0.448	0.483	-7.8	97	0.00
73	Bromobenzene	0.436	0.409	6.2	98	0.00
74	n-propylbenzene	2.174	2.177	-0.1	98	0.00
75	2-Chlorotoluene	1.305	1.327	-1.7	99	0.00
76	1,3,5-Trimethylbenzene	1.562	1.574	-0.8	99	0.00
77	t-1,4-Dichloro-2-butene	0.151	0.136	9.9	100	0.00
78	4-Chlorotoluene	1.301	1.292	0.7	96	0.00
79	tert-butylbenzene	1.341	1.291	3.7	95	0.00
80	1,2,4-Trimethylbenzene	1.551	1.540	0.7	97	0.00
81	sec-Butylbenzene	1.860	1.809	2.7	94	0.00
82	p-Isopropyltoluene	1.548	1.450	6.3	95	0.00
83 M	1,3-Dichlorobenzene	0.814	0.768	5.7	97	0.00
84 M	1,4-Dichlorobenzene	0.793	0.738	6.9	95	0.00
85	n-Butylbenzene	1.285	1.189	7.5	93	0.00
86 T	Hexachloroethane	0.258	0.243	5.8	100	0.00
87 M	1,2-Dichlorobenzene	0.777	0.741	4.6	97	0.00
88	1,2-Dibromo-3-Chloropropane	0.096	0.095	1.0	100	0.00
89	1,2,4-Trichlorobenzene	0.354	0.285	19.5	90	0.00
90	Hexachlorobutadiene	0.170	0.153	10.0	87	0.00
91 M	Naphthalene	1.007	0.754	25.1	88	0.00
92	1,2,3-Trichlorobenzene	0.340	0.283	16.8	94	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0