

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072225\
 Data File : VN087392.D
 Acq On : 22 Jul 2025 16:12
 Operator : JC\MD
 Sample : VN0722WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0722WBL01

Quant Time: Jul 23 05:56:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N062525W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Jun 25 10:49:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.794	128	27646	30.000	ug/l	-0.03
28) 1,4-Difluorobenzene	9.088	114	129340	30.000	ug/l	-0.02
57) Chlorobenzene-d5	11.847	117	109767	30.000	ug/l	-0.02

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.565	65	65963	31.667	ug/l	-0.02
Spiked Amount	30.000	Range	91 - 110	Recovery	=	105.567%
60) 4-Bromofluorobenzene	12.829	95	46084	27.207	ug/l	-0.02
Spiked Amount	30.000	Range	63 - 112	Recovery	=	90.700%
63) Toluene-d8	10.547	98	149028	30.238	ug/l	-0.02
Spiked Amount	30.000	Range	91 - 112	Recovery	=	100.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072225\
Data File : VN087392.D
Acq On : 22 Jul 2025 16:12
Operator : JC\MD
Sample : VN0722WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0722WBL01

Quant Time: Jul 23 05:56:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N062525W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Jun 25 10:49:56 2025
Response via : Initial Calibration

