

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072318\
 Data File : VN050004.D
 Acq On : 23 Jul 2018 11:11
 Operator : MD\SY
 Sample : PB111341TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB111341TB

Quant Time: Jul 24 05:05:36 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 19 01:58:39 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	625930	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	993529	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	860817	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	302229	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	432927	50.32	ug/l	0.00
Spiked Amount						
			Recovery	=	100.64%	
35) Dibromofluoromethane	7.59	113	388443	46.57	ug/l	0.00
Spiked Amount						
			Recovery	=	93.14%	
50) Toluene-d8	10.09	98	1361891	44.81	ug/l	0.00
Spiked Amount						
			Recovery	=	89.62%	
62) 4-Bromofluorobenzene	12.40	95	371852	35.53	ug/l	0.00
Spiked Amount						
			Recovery	=	71.06%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.83	43	24303	10.91	ug/l	92
18) Methyl Acetate	4.34	43	18289	1.82	ug/l	95
20) Methylene Chloride	4.54	84	15367	1.81	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN072318\
Data File : VN050004.D
Acq On : 23 Jul 2018 11:11
Operator : MD\SY
Sample : PB111341TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB111341TB

Quant Time: Jul 24 05:05:36 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 19 01:58:39 2018
Response via : Initial Calibration

