

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072318\
 Data File : VN050006.D
 Acq On : 23 Jul 2018 12:03
 Operator : MD\SY
 Sample : PB111341ZHE#02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB111341ZHE#02

Quant Time: Jul 24 05:15:15 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 19 01:58:39 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	605406	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	999429	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	836356	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	244910	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	428349	51.48	ug/l	0.00
Spiked Amount						
			Recovery	=	102.96%	
35) Dibromofluoromethane	7.59	113	376775	44.91	ug/l	0.00
Spiked Amount						
			Recovery	=	89.82%	
50) Toluene-d8	10.09	98	1358597	44.44	ug/l	0.00
Spiked Amount						
			Recovery	=	88.88%	
62) 4-Bromofluorobenzene	12.40	95	341660	32.45	ug/l	0.00
Spiked Amount						
			Recovery	=	64.90%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.83	43	25139	11.67	ug/l	100
18) Methyl Acetate	4.33	43	20920	2.30	ug/l #	71
20) Methylene Chloride	4.55	84	15231	1.86	ug/l	93
25) 2-Butanone	6.85	43	33284	10.81	ug/l	96
29) Tetrahydrofuran	7.22	42	16283	8.39	ug/l #	82
43) Isopropyl Acetate	8.17	43	280828	21.06	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN072318\
Data File : VN050006.D
Acq On : 23 Jul 2018 12:03
Operator : MD\SY
Sample : PB111341ZHE#02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB111341ZHE#02

Quant Time: Jul 24 05:15:15 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 19 01:58:39 2018
Response via : Initial Calibration

