

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072318\
 Data File : VN050008.D
 Acq On : 23 Jul 2018 12:55
 Operator : MD\SY
 Sample : J4101-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-2

Quant Time: Jul 24 05:38:04 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 19 01:58:39 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	586420	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	991028	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	856944	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	325471	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	425685	52.81	ug/l	0.00
Spiked Amount						
			Recovery	=	105.62%	
35) Dibromofluoromethane	7.59	113	371209	44.62	ug/l	0.00
Spiked Amount						
			Recovery	=	89.24%	
50) Toluene-d8	10.09	98	1353991	44.67	ug/l	0.00
Spiked Amount						
			Recovery	=	89.34%	
62) 4-Bromofluorobenzene	12.40	95	381727	36.56	ug/l	0.00
Spiked Amount						
			Recovery	=	73.12%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.83	43	17248	8.27	ug/l	99
18) Methyl Acetate	4.33	43	26884	3.33	ug/l	94
20) Methylene Chloride	4.55	84	603544	76.05	ug/l	99
29) Tetrahydrofuran	7.22	42	3752	2.00	ug/l #	83
30) Chloroform	7.37	83	8986	0.67	ug/l	94
43) Isopropyl Acetate	8.17	43	414715	31.37	ug/l #	87
47) Bromodichloromethane	9.41	83	3084	0.27	ug/l	90
95) Naphthalene	15.14	128	411472	28.79	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN072318\
Data File : VN050008.D
Acq On : 23 Jul 2018 12:55
Operator : MD\SY
Sample : J4101-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-2

Quant Time: Jul 24 05:38:04 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 19 01:58:39 2018
Response via : Initial Calibration

