


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN072318\
 Data File : VN049999.D
 Acq On : 23 Jul 2018 7:49
 Operator : MD\SY
 Sample : BFB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
 Title : SW846 8260
 Last Update : Thu Jul 19 01:58:39 2018



AutoFind: Scans 3362, 3363, 3364; Background Corrected with Scan 3348

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.1	10292	PASS
75	95	30	60	47.4	25544	PASS
95	95	100	100	100.0	53928	PASS
96	95	5	9	7.2	3895	PASS
173	174	0.00	2	1.2	625	PASS
174	95	50	100	93.2	50240	PASS
175	174	5	9	7.0	3534	PASS
176	174	95	101	95.4	47949	PASS
177	176	5	9	6.8	3279	PASS