

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072319\
 Data File : VN056873.D
 Acq On : 23 Jul 2019 12:59
 Operator : JC/SP
 Sample : K3617-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 NB-07-071919-B

Quant Time: Jul 24 01:43:34 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072219W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 23 01:46:34 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	242469	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	448622	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	499926	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	174097	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	144617	55.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.84%	
35) Dibromofluoromethane	7.59	113	138117	48.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.14%	
50) Toluene-d8	10.09	98	573936	52.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.52%	
62) 4-Bromofluorobenzene	12.40	95	211114	59.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.60%	
Target Compounds						
16) Acetone	3.82	43	14621	15.699	ug/l	93
20) Methylene Chloride	4.54	84	61660	22.141	ug/l	98
30) Chloroform	7.37	83	5731	1.262	ug/l	89
43) Isopropyl Acetate	8.17	43	15134	2.639	ug/l	94
95) Naphthalene	15.13	128	49801	9.078	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN072319\
Data File : VN056873.D
Acq On : 23 Jul 2019 12:59
Operator : JC/SP
Sample : K3617-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
NB-07-071919-B

Quant Time: Jul 24 01:43:34 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072219W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 23 01:46:34 2019
Response via : Initial Calibration

