

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072319\
 Data File : VN056875.D
 Acq On : 23 Jul 2019 13:47
 Operator : JC/SP
 Sample : K3622-10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OR-02-071919-A

Quant Time: Jul 24 01:48:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072219W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 23 01:46:34 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	274109	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	499590	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	570105	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	203711	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	161671	55.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.60%	
35) Dibromofluoromethane	7.59	113	151893	47.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.94%	
50) Toluene-d8	10.09	98	654706	54.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.08%	
62) 4-Bromofluorobenzene	12.40	95	242998	61.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	122.58%	
Target Compounds						
16) Acetone	3.82	43	13889	13.191	ug/l	92
20) Methylene Chloride	4.54	84	210739	66.938	ug/l	99
30) Chloroform	7.37	83	7377	1.437	ug/l	99
43) Isopropyl Acetate	8.17	43	15714	2.460	ug/l	98
95) Naphthalene	15.13	128	14345	4.968	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN072319\
Data File : VN056875.D
Acq On : 23 Jul 2019 13:47
Operator : JC/SP
Sample : K3622-10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OR-02-071919-A

Quant Time: Jul 24 01:48:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072219W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 23 01:46:34 2019
Response via : Initial Calibration

