

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072319\
 Data File : VN056876.D
 Acq On : 23 Jul 2019 14:12
 Operator : JC/SP
 Sample : K3621-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OK-02-071919

Quant Time: Jul 24 01:52:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072219W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 23 01:46:34 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	242351	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	446638	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	504701	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	167202	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	143837	55.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.28%	
35) Dibromofluoromethane	7.59	113	133282	47.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.16%	
50) Toluene-d8	10.09	98	574208	53.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.04%	
62) 4-Bromofluorobenzene	12.40	95	211266	59.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	119.22%	
Target Compounds						
16) Acetone	3.82	43	12377	13.296	ug/l	94
20) Methylene Chloride	4.55	84	423616	152.188	ug/l	98
43) Isopropyl Acetate	8.17	43	15954	2.794	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072319\
Data File : VN056876.D
Acq On : 23 Jul 2019 14:12
Operator : JC/SP
Sample : K3621-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OK-02-071919

Quant Time: Jul 24 01:52:11 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072219W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 23 01:46:34 2019
Response via : Initial Calibration

