

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072319\
 Data File : VN056881.D
 Acq On : 23 Jul 2019 16:14
 Operator : JC/SP
 Sample : PB121628ZHE#02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB121628ZHE#02

Quant Time: Jul 24 02:04:47 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072219W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 23 01:46:34 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	236200	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	449443	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	501963	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	175821	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	142157	56.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.84%	
35) Dibromofluoromethane	7.59	113	134840	47.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.66%	
50) Toluene-d8	10.09	98	574998	52.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.52%	
62) 4-Bromofluorobenzene	12.40	95	214728	60.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	120.40%	
Target Compounds						
16) Acetone	3.82	43	24139	26.606	ug/l	96
18) Methyl Acetate	4.34	43	13145	2.681	ug/l	93
20) Methylene Chloride	4.54	84	4567	1.683	ug/l	96
43) Isopropyl Acetate	8.17	43	17864	3.109	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN072319\
Data File : VN056881.D
Acq On : 23 Jul 2019 16:14
Operator : JC/SP
Sample : PB121628ZHE#02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB121628ZHE#02

Quant Time: Jul 24 02:04:47 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072219W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 23 01:46:34 2019
Response via : Initial Calibration

