

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072319\
 Data File : VN056883.D
 Acq On : 23 Jul 2019 17:02
 Operator : JC/SP
 Sample : K3943-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PA-71619

Quant Time: Jul 24 02:20:40 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072219W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 23 01:46:34 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	236074	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	447211	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	500437	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	169228	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	144257	57.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.58%	
35) Dibromofluoromethane	7.59	113	133135	46.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.94%	
50) Toluene-d8	10.09	98	573137	52.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.70%	
62) 4-Bromofluorobenzene	12.40	95	210325	59.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.52%	
Target Compounds						
16) Acetone	3.82	43	16143	17.803	ug/l	97
18) Methyl Acetate	4.33	43	15731	3.586	ug/l	97
20) Methylene Chloride	4.55	84	95495	35.220	ug/l	99
43) Isopropyl Acetate	8.17	43	15815	2.766	ug/l	97
95) Naphthalene	15.13	128	14694	5.281	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN072319\
Data File : VN056883.D
Acq On : 23 Jul 2019 17:02
Operator : JC/SP
Sample : K3943-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PA-71619

Quant Time: Jul 24 02:20:40 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072219W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 23 01:46:34 2019
Response via : Initial Calibration

