

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072321\
 Data File : VN067884.D
 Acq On : 23 Jul 2021 16:53
 Operator : JC/MD
 Sample : M3147-05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 152140-MW-9(ONSITE)

Manual Integrations
 APPROVED

MMDadoda
 7/26/2021 5:21:55 PM

Quant Time: Jul 24 01:42:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 07 16:18:53 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	276626	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	538856	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	505933	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	171028	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	224519	53.704	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery	= 107.400%		
35) Dibromofluoromethane	8.024	113	164325	50.474	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery	= 100.940%		
50) Toluene-d8	10.446	98	677624	49.726	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery	= 99.460%		
62) 4-Bromofluorobenzene	12.734	95	216396	44.952	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery	= 89.900%		
Target Compounds						Qvalue
16) Acetone	4.307	43	4769m	3.625	ug/l	
44) Trichloroethene	9.217	130	3389	0.891	ug/l	73
64) Tetrachloroethene	10.980	164	2240	0.627	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072321\
Data File : VN067884.D
Acq On : 23 Jul 2021 16:53
Operator : JC/MD
Sample : M3147-05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
152140-MW-9(ONSITE)

Manual Integrations
APPROVED

MMDadoda
7/26/2021 5:21:55 PM

Quant Time: Jul 24 01:42:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 07 16:18:53 2021
Response via : Initial Calibration

