

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072321\
 Data File : VN067891.D
 Acq On : 23 Jul 2021 19:45
 Operator : JC/MD
 Sample : M3150-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-B

Manual Integrations
 APPROVED

MMDadoda
 7/26/2021 5:22:14 PM

Quant Time: Jul 24 01:43:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 07 16:18:53 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	281083	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	548993	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	565563	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	194681	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	228184	53.715	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 107.440%	
35) Dibromofluoromethane	8.024	113	168036	50.661	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 101.320%	
50) Toluene-d8	10.446	98	716280	51.592	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 103.180%	
62) 4-Bromofluorobenzene	12.736	95	241451	49.231	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 98.460%	
Target Compounds						
					Qvalue	
16) Acetone	4.296	43	20431	15.285	ug/l	99
18) Methyl Acetate	4.862	43	11408m	3.039	ug/l	
20) Methylene Chloride	5.092	84	25626	7.623	ug/l	88
43) Isopropyl Acetate	8.566	43	39836	4.583	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072321\
Data File : VN067891.D
Acq On : 23 Jul 2021 19:45
Operator : JC/MD
Sample : M3150-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-B

Manual Integrations
APPROVED

MMDadoda
7/26/2021 5:22:14 PM

Quant Time: Jul 24 01:43:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 07 16:18:53 2021
Response via : Initial Calibration

