

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072321\
 Data File : VN067892.D
 Acq On : 23 Jul 2021 20:09
 Operator : JC/MD
 Sample : M3150-05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-C

Quant Time: Jul 24 01:44:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 07 16:18:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	286079	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	567472	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	591555	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	210221	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	230015	53.201	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	106.400%
35) Dibromofluoromethane	8.024	113	170443	49.713	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.420%
50) Toluene-d8	10.443	98	734949	51.213	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.420%
62) 4-Bromofluorobenzene	12.734	95	250498	49.412	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	98.820%
Target Compounds						
					Qvalue	
16) Acetone	4.296	43	34853	25.619	ug/l	94
18) Methyl Acetate	4.875	43	10959	2.868	ug/l	96
20) Methylene Chloride	5.101	84	25551	7.453	ug/l #	83
43) Isopropyl Acetate	8.566	43	39951	4.447	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072321\
Data File : VN067892.D
Acq On : 23 Jul 2021 20:09
Operator : JC/MD
Sample : M3150-05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-C

Quant Time: Jul 24 01:44:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 07 16:18:53 2021
Response via : Initial Calibration

