

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072324\
 Data File : VN083011.D
 Acq On : 23 Jul 2024 13:37
 Operator : JC\MD
 Sample : VN0723MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0723MBL01

Quant Time: Jul 24 02:28:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 13:46:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	57076	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	108408	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	110286	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	42709	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	51601	64.494	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 128.980%#	
35) Dibromofluoromethane	8.171	113	42932	61.717	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 123.440%	
50) Toluene-d8	10.571	98	145680	56.701	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 113.400%#	
62) 4-Bromofluorobenzene	12.853	95	42966	46.555	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 93.100%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072324\
Data File : VN083011.D
Acq On : 23 Jul 2024 13:37
Operator : JC\MD
Sample : VN0723MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0723MBL01

Quant Time: Jul 24 02:28:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 08 13:46:04 2024
Response via : Initial Calibration

