

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072324\
 Data File : VN083014.D
 Acq On : 23 Jul 2024 16:23
 Operator : JC\MD
 Sample : P3302-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 285-290

Quant Time: Jul 23 16:41:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 13:46:04 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	66762	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	131528	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	145057	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	54576	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	52034	55.600	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.200%
35) Dibromofluoromethane	8.171	113	43878	51.989	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.980%
50) Toluene-d8	10.571	98	165549	53.108	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	106.220%
62) 4-Bromofluorobenzene	12.853	95	49117	43.865	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	87.720%
Target Compounds						
						Qvalue
16) Acetone	4.436	43	10140	30.774	ug/l	93
20) Methylene Chloride	5.277	84	6762	8.498	ug/l #	69
43) Isopropyl Acetate	8.694	43	59414	25.121	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072324\
Data File : VN083014.D
Acq On : 23 Jul 2024 16:23
Operator : JC\MD
Sample : P3302-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
285-290

Quant Time: Jul 23 16:41:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 08 13:46:04 2024
Response via : Initial Calibration

