

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072324\
 Data File : VN083015.D
 Acq On : 23 Jul 2024 16:47
 Operator : JC\MD
 Sample : PB162199TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB162199TB

Quant Time: Jul 24 02:30:16 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 13:46:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	66960	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	136583	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	151986	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	59119	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	54169	57.710	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	115.420%
35) Dibromofluoromethane	8.171	113	44492	50.766	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.540%
50) Toluene-d8	10.571	98	169246	52.284	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.560%
62) 4-Bromofluorobenzene	12.853	95	52144	44.845	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	89.680%
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	8646	26.163	ug/l	97
20) Methylene Chloride	5.277	84	3074	3.852	ug/l #	79
25) 2-Butanone	7.488	43	3687	7.389	ug/l	92
43) Isopropyl Acetate	8.688	43	74189	30.846	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072324\
Data File : VN083015.D
Acq On : 23 Jul 2024 16:47
Operator : JC\MD
Sample : PB162199TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB162199TB

Quant Time: Jul 24 02:30:16 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 08 13:46:04 2024
Response via : Initial Calibration

