

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072324\
 Data File : VN083016.D
 Acq On : 23 Jul 2024 17:12
 Operator : JC\MD
 Sample : PB162199ZHE#01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB162199ZHE#01

Quant Time: Jul 24 02:30:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 13:46:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	65188	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	136542	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	149388	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	56526	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	53140	58.153	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	116.300%
35) Dibromofluoromethane	8.171	113	44619	50.926	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.860%
50) Toluene-d8	10.570	98	170111	52.567	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	105.140%
62) 4-Bromofluorobenzene	12.853	95	52673	45.313	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	90.620%
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	8611	26.765	ug/l	93
20) Methylene Chloride	5.271	84	4294	5.527	ug/l #	78
25) 2-Butanone	7.494	43	3715	7.647	ug/l	95
43) Isopropyl Acetate	8.688	43	79450	33.268	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072324\
Data File : VN083016.D
Acq On : 23 Jul 2024 17:12
Operator : JC\MD
Sample : PB162199ZHE#01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB162199ZHE#01

Quant Time: Jul 24 02:30:49 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 08 13:46:04 2024
Response via : Initial Calibration

