


Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072418\
Data File : VN050009.D
Acq On : 23 Jul 2018 17:01
Operator : MD\SY
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
Title : SW846 8260
Last Update : Thu Jul 26 18:10:10 2018



AutoFind: Scans 3362, 3363, 3364; Background Corrected with Scan 3349

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.3	8840	PASS
75	95	30	60	51.9	25045	PASS
95	95	100	100	100.0	48253	PASS
96	95	5	9	7.2	3454	PASS
173	174	0.00	2	2.0	924	PASS
174	95	50	100	96.3	46469	PASS
175	174	5	9	7.1	3305	PASS
176	174	95	101	95.0	44160	PASS
177	176	5	9	7.2	3186	PASS