

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072419\
 Data File : VN056898.D
 Acq On : 24 Jul 2019 1:06
 Operator : JC/SP
 Sample : K3908-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PWU-3

Quant Time: Jul 24 04:13:40 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N072419W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Jul 24 03:39:51 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.19	128	54883	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.59	114	283101	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	220427	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	8.03	65	109147	30.39	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	101.30%
60) 4-Bromofluorobenzene	12.40	95	71469	21.63	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	72.10%
63) Toluene-d8	10.09	98	336172	32.29	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	107.63%

Target Compounds

					Ovalue	
24) Methyl tert-Butyl Ether	5.05	73	16756	1.892	ug/l #	75
91) Naphthalene	15.13	128	3195	8.058	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN072419\
Data File : VN056898.D
Acq On : 24 Jul 2019 1:06
Operator : JC/SP
Sample : K3908-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PWU-3

Quant Time: Jul 24 04:13:40 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\624N072419W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Jul 24 03:39:51 2019
Response via : Initial Calibration

