

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072423\
 Data File : VN078637.D
 Acq On : 24 Jul 2023 14:52
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Jul 25 03:14:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 10 16:13:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	476927	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	833608	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	756050	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	280245	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	291403	45.576	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	91.160%
35) Dibromofluoromethane	8.171	113	291972	51.596	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.200%
50) Toluene-d8	10.577	98	988266	46.230	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.460%
62) 4-Bromofluorobenzene	12.853	95	337546	51.263	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.520%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072423\
Data File : VN078637.D
Acq On : 24 Jul 2023 14:52
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Jul 25 03:14:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 10 16:13:53 2023
Response via : Initial Calibration

