

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072424\
 Data File : VN083032.D
 Acq On : 24 Jul 2024 16:31
 Operator : JC\MD
 Sample : P3338-04RE
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Jul 25 02:41:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 13:46:04 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	8.230	168	144524	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	284110	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	282930	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	109504	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	113312	55.931	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 111.860%	
35) Dibromofluoromethane	8.165	113	97277	53.359	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.720%	
50) Toluene-d8	10.571	98	327102	48.579	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 97.160%	
62) 4-Bromofluorobenzene	12.853	95	102654	42.442	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 84.880%	
Target Compounds						
					Qvalue	
3) Chloromethane	2.365	50	786	0.443	ug/l	93
5) Bromomethane	2.959	94	962	0.716	ug/l #	73
16) Acetone	4.442	43	3250	4.556	ug/l #	83
31) Cyclohexane	8.230	56	4121	1.556	ug/l #	7
36) 1,1-Dichloropropene	8.224	75	13772	5.479	ug/l #	51
38) Carbon Tetrachloride	8.230	117	15419	5.095	ug/l #	18
43) Isopropyl Acetate	8.694	43	1446	Below Cal	#	60
71) Bromoform	12.853	173	1265	0.746	ug/l #	1
76) 1,2,3-Trichloropropane	12.853	75	60039	31.802	ug/l #	1
81) trans-1,4-Dichloro-2-b...	12.853	75	60039	62.545	ug/l #	6


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072424\
Data File : VN083032.D
Acq On : 24 Jul 2024 16:31
Operator : JC\MD
Sample : P3338-04RE
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :

Quant Time: Jul 25 02:41:02 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 08 13:46:04 2024
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.230 min Scan# 1
Delta R.T. 0.001 min
Lab File: VN083032.D
Acq: 24 Jul 2024 16:31




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:168 Resp: 144524
Ion Ratio Lower Upper
168 100
99 53.5 48.2 72.4



#3
Chloromethane
Concen: 0.443 ug/l
RT: 2.365 min Scan# 70
Delta R.T. 0.006 min
Lab File: VN083032.D
Acq: 24 Jul 2024 16:31

Tgt Ion: 50 Resp: 786
Ion Ratio Lower Upper
50 100
52 35.1 25.0 37.4









#5
Bromomethane
Concen: 0.716 ug/l
RT: 2.959 min Scan# 11
Delta R.T. 0.012 min
Lab File: VN083032.D
Acq: 24 Jul 2024 16:31




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 94 Resp: 962
Ion Ratio Lower Upper
94 100
96 71.1 78.0 117.0#



#16
Acetone
Concen: 4.556 ug/l
RT: 4.442 min Scan# 423
Delta R.T. 0.012 min
Lab File: VN083032.D
Acq: 24 Jul 2024 16:31


Tgt Ion: 43 Resp: 3250
Ion Ratio Lower Upper
43 100
58 36.2 21.8 32.6#








#31
Cyclohexane
Concen: 1.556 ug/l
RT: 8.230 min Scan# 1067
Delta R.T. -0.029 min
Lab File: VN083032.D
Acq: 24 Jul 2024 16:31

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 56 Resp: 4121
Ion Ratio Lower Upper
56 100
69 135.5 24.5 36.7#
84 133.5 66.0 99.0#



#33
1,2-Dichloroethane-d4
Concen: 55.931 ug/l
RT: 8.583 min Scan# 1127
Delta R.T. 0.001 min
Lab File: VN083032.D
Acq: 24 Jul 2024 16:31

Tgt Ion: 65 Resp: 113312
Ion Ratio Lower Upper
65 100
67 51.2 0.0 104.4









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 9.106 min Scan# 11
 Delta R.T. 0.000 min
 Lab File: VN083032.D
 Acq: 24 Jul 2024 16:31




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion:114 Resp: 284110
 Ion Ratio Lower Upper
 114 100
 63 21.2 0.0 44.6
 88 14.6 0.0 31.4



#35
 Dibromofluoromethane
 Concen: 53.359 ug/l
 RT: 8.165 min Scan# 1056
 Delta R.T. -0.000 min
 Lab File: VN083032.D
 Acq: 24 Jul 2024 16:31

Tgt Ion:113 Resp: 97277
 Ion Ratio Lower Upper
 113 100
 111 101.5 82.4 123.6
 192 17.8 14.9 22.3





#36

1,1-Dichloropropene

Concen: 5.479 ug/l

RT: 8.224 min Scan# 1066

Delta R.T. -0.147 min

Lab File: VN083032.D





Acq: 24 Jul 2024 16:31

Instrument :

MSVOA_N

ClientSampleId :

Tgt Ion:	75	Resp:	13772
Ion	Ratio	Lower	Upper
75	100		
110	8.9	16.3	48.9#
77	0.0	24.6	37.0#



#38

Carbon Tetrachloride

Concen: 5.095 ug/l




RT: 8.230 min Scan# 1067


Delta R.T. -0.135 min

Lab File: VN083032.D

Acq: 24 Jul 2024 16:31

Tgt Ion:	117	Resp:	15419
Ion	Ratio	Lower	Upper
117	100		
119	6.2	74.9	112.3#
121	0.0	24.3	36.5#





#43

Isopropyl Acetate

Concen: Below Cal

RT: 8.694 min Scan# 1145

Delta R.T. 0.000 min

Lab File: VN083032.D

Acq: 24 Jul 2024 16:31

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion: 43 Resp: 1446

Ion Ratio Lower Upper

43 100

61 0.0 17.8 26.6#

87 0.0 8.2 12.2#



#50

Toluene-d8

Concen: 48.579 ug/l

RT: 10.571 min Scan# 1465

Delta R.T. 0.001 min

Lab File: VN083032.D




Acq: 24 Jul 2024 16:31


Tgt Ion: 98 Resp: 327102

Ion Ratio Lower Upper

98 100

100 61.9 51.5 77.3









#62
4-Bromofluorobenzene
Concen: 42.442 ug/l
RT: 12.853 min Scan# 1853
Delta R.T. 0.006 min
Lab File: VN083032.D
Acq: 24 Jul 2024 16:31




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 95 Resp: 102654
Ion Ratio Lower Upper
95 100
174 82.8 0.0 159.2
176 79.4 0.0 147.6



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.871 min Scan# 1686
Delta R.T. 0.001 min
Lab File: VN083032.D
Acq: 24 Jul 2024 16:31

Tgt Ion: 117 Resp: 282930
Ion Ratio Lower Upper
117 100
82 53.6 47.5 71.3
119 31.5 25.6 38.4





#71

Bromoform

Concen: 0.746 ug/l

RT: 12.853 min Scan# 11

Delta R.T. 0.271 min


Lab File: VN083032.D

Acq: 24 Jul 2024 16:31




Instrument :

MSVOA_N

ClientSampleId :



Tgt Ion:	173	Resp:	1265
Ion Ratio	Lower	Upper	
173	100		
175	463.0	24.0	72.0#
254	0.0	0.0	0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 13.794 min Scan# 2013


Delta R.T. 0.000 min

Lab File: VN083032.D

Acq: 24 Jul 2024 16:31

Tgt Ion:	152	Resp:	109504
Ion Ratio	Lower	Upper	
152	100		
115	58.9	30.6	91.6
150	153.4	0.0	348.6









#76
1,2,3-Trichloropropane
Concen: 31.802 ug/l
RT: 12.853 min Scan# 1853
Delta R.T. -0.141 min
Lab File: VN083032.D
Acq: 24 Jul 2024 16:31

Instrument :
MSVOA_N
ClientSampleId :

Tgt Ion: 75 Resp: 60039
Ion Ratio Lower Upper
75 100
77 0.8 110.9 332.6#



#81
trans-1,4-Dichloro-2-butene
Concen: 62.545 ug/l
RT: 12.853 min Scan# 1853
Delta R.T. 0.112 min
Lab File: VN083032.D
Acq: 24 Jul 2024 16:31

Tgt Ion: 75 Resp: 60039
Ion Ratio Lower Upper
75 100
53 0.3 88.4 132.6#
89 0.0 36.3 54.5#

