

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072424\
 Data File : VN083027.D
 Acq On : 24 Jul 2024 14:29
 Operator : JC\MD
 Sample : P3338-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-196-EB-20240718

Quant Time: Jul 25 02:38:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 13:46:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	152186	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	294803	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	290676	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	116689	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	120869	56.657	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 113.320%			
35) Dibromofluoromethane	8.171	113	101241	53.519	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 107.040%			
50) Toluene-d8	10.570	98	340196	48.691	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 97.380%			
62) 4-Bromofluorobenzene	12.847	95	113162	45.089	ug/l	0.00
Spiked Amount 50.000	Range 64 - 133		Recovery = 90.180%			
Target Compounds						
16) Acetone	4.436	43	14853	19.775	ug/l	Qvalue # 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072424\
Data File : VN083027.D
Acq On : 24 Jul 2024 14:29
Operator : JC\MD
Sample : P3338-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-196-EB-20240718

Quant Time: Jul 25 02:38:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 08 13:46:04 2024
Response via : Initial Calibration

