

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072424\
 Data File : VN083029.D
 Acq On : 24 Jul 2024 15:18
 Operator : JC\MD
 Sample : P3338-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-196-GW-1000-1002

Quant Time: Jul 25 02:39:42 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 13:46:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	141249	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	284089	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	289464	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	115260	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	117725	59.456	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	118.920%	
35) Dibromofluoromethane	8.171	113	98222	53.881	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	107.760%	
50) Toluene-d8	10.571	98	332447	49.376	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	98.760%	
62) 4-Bromofluorobenzene	12.847	95	110032	45.495	ug/l	0.00
Spiked Amount	50.000	Range 64 - 133	Recovery	=	91.000%	
Target Compounds						
16) Acetone	4.442	43	3694	5.299	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072424\
Data File : VN083029.D
Acq On : 24 Jul 2024 15:18
Operator : JC\MD
Sample : P3338-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-196-GW-1000-1002

Quant Time: Jul 25 02:39:42 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 08 13:46:04 2024
Response via : Initial Calibration

