

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072424\
 Data File : VN083030.D
 Acq On : 24 Jul 2024 15:42
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Jul 25 02:40:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 13:46:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	146463	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	288321	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	300026	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	120911	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	118674	57.802	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	115.600%	
35) Dibromofluoromethane	8.171	113	100366	54.249	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	108.500%	
50) Toluene-d8	10.571	98	343857	50.321	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	100.640%	
62) 4-Bromofluorobenzene	12.853	95	113376	46.190	ug/l	0.00
Spiked Amount	50.000	Range 64 - 133	Recovery	=	92.380%	

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072424\
Data File : VN083030.D
Acq On : 24 Jul 2024 15:42
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Jul 25 02:40:10 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 08 13:46:04 2024
Response via : Initial Calibration

