

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072518\
 Data File : VN050042.D
 Acq On : 25 Jul 2018 11:25
 Operator : MD\SY
 Sample : VN0725WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0725WBL01

Quant Time: Jul 26 06:41:02 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N072518W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Jul 25 10:32:46 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.20	128	128579	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.59	114	676471	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	549628	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.03	65	271060	31.01	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	103.37%
60) 4-Bromofluorobenzene	12.40	95	193564	22.92	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	76.40%
63) Toluene-d8	10.09	98	781476	29.77	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	99.23%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072518\
Data File : VN050042.D
Acq On : 25 Jul 2018 11:25
Operator : MD\SY
Sample : VN0725WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0725WBL01

Quant Time: Jul 26 06:41:02 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N072518W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Jul 25 10:32:46 2018
Response via : Initial Calibration

