

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072518\
 Data File : VN050058.D
 Acq On : 25 Jul 2018 21:50
 Operator : MD\SY
 Sample : PB111424ZHE#05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB111424ZHE#05

Quant Time: Jul 26 07:25:29 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 24 06:27:04 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	616780	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1030137	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	848569	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	263907	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	441440	51.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.60%	
35) Dibromofluoromethane	7.59	113	393261	46.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.52%	
50) Toluene-d8	10.09	98	1381031	44.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.32%	
62) 4-Bromofluorobenzene	12.41	95	352925	34.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	69.32%	
Target Compounds						
16) Acetone	3.83	43	23688	7.47	ug/l	94
18) Methyl Acetate	4.33	43	22868	1.31	ug/l	97
43) Isopropyl Acetate	8.17	43	292098	18.44	ug/l #	90
59) 2-Hexanone	10.77	43	16873	8.00	ug/l #	62

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN072518\
Data File : VN050058.D
Acq On : 25 Jul 2018 21:50
Operator : MD\SY
Sample : PB111424ZHE#05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB111424ZHE#05

Quant Time: Jul 26 07:25:29 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 24 06:27:04 2018
Response via : Initial Calibration

