

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072518\
 Data File : VN050059.D
 Acq On : 25 Jul 2018 22:14
 Operator : MD\SY
 Sample : PB111424ZHE#06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB111424ZHE#06

Quant Time: Jul 26 07:28:12 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 24 06:27:04 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	608442	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1026322	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	880551	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	309833	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	443943	52.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.60%	
35) Dibromofluoromethane	7.60	113	394576	47.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.18%	
50) Toluene-d8	10.09	98	1389287	45.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.20%	
62) 4-Bromofluorobenzene	12.40	95	381246	37.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.16%	
Target Compounds						
16) Acetone	3.83	43	23964	7.78	ug/l	98
18) Methyl Acetate	4.34	43	24260	1.54	ug/l	97
43) Isopropyl Acetate	8.17	43	309994	19.71	ug/l #	89
59) 2-Hexanone	10.77	43	17984	8.20	ug/l #	60

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN072518\
Data File : VN050059.D
Acq On : 25 Jul 2018 22:14
Operator : MD\SY
Sample : PB111424ZHE#06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB111424ZHE#06

Quant Time: Jul 26 07:28:12 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 24 06:27:04 2018
Response via : Initial Calibration

