

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072518\
 Data File : VN050061.D
 Acq On : 25 Jul 2018 23:04
 Operator : MD\SY
 Sample : PB111424ZHE#08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB111424ZHE#08

Quant Time: Jul 26 07:32:17 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 24 06:27:04 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	601445	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1024910	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	869180	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	290688	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	439198	52.34	ug/l	0.00
Spiked Amount						
			Recovery	=	104.68%	
35) Dibromofluoromethane	7.59	113	390501	46.67	ug/l	0.00
Spiked Amount						
			Recovery	=	93.34%	
50) Toluene-d8	10.09	98	1390452	45.20	ug/l	0.00
Spiked Amount						
			Recovery	=	90.40%	
62) 4-Bromofluorobenzene	12.40	95	366993	36.23	ug/l	0.00
Spiked Amount						
			Recovery	=	72.46%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	22472	7.15	ug/l	99
18) Methyl Acetate	4.33	43	25269	1.71	ug/l	100
43) Isopropyl Acetate	8.17	43	354657	22.76	ug/l #	88
59) 2-Hexanone	10.77	43	18157	8.23	ug/l #	61

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN072518\
Data File : VN050061.D
Acq On : 25 Jul 2018 23:04
Operator : MD\SY
Sample : PB111424ZHE#08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB111424ZHE#08

Quant Time: Jul 26 07:32:17 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 24 06:27:04 2018
Response via : Initial Calibration

