

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072518\
 Data File : VN050064.D
 Acq On : 26 Jul 2018 00:19
 Operator : MD\SY
 Sample : J4135-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CON-WC-S-2-0-11

Quant Time: Jul 26 07:45:41 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 24 06:27:04 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	595000	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1004461	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	860066	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	369167	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	432087	52.05	ug/l	0.00
Spiked Amount						
			Recovery	=	104.10%	
35) Dibromofluoromethane	7.59	113	389176	47.45	ug/l	0.00
Spiked Amount						
			Recovery	=	94.90%	
50) Toluene-d8	10.09	98	1360035	45.11	ug/l	0.00
Spiked Amount						
			Recovery	=	90.22%	
62) 4-Bromofluorobenzene	12.40	95	389693	39.25	ug/l	0.00
Spiked Amount						
			Recovery	=	78.50%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	28393	10.36	ug/l	99
18) Methyl Acetate	4.33	43	27293	2.03	ug/l	96
20) Methylene Chloride	4.55	84	596334	81.82	ug/l	99
43) Isopropyl Acetate	8.17	43	296886	19.27	ug/l #	89
44) Trichloroethene	8.84	130	12724	1.38	ug/l	95
59) 2-Hexanone	10.77	43	15012	7.75	ug/l	63
95) Naphthalene	15.14	128	12299088	669.50	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072518\
Data File : VN050064.D
Acq On : 26 Jul 2018 00:19
Operator : MD\SY
Sample : J4135-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CON-WC-S-2-0-11

Quant Time: Jul 26 07:45:41 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 24 06:27:04 2018
Response via : Initial Calibration

