

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072518\
 Data File : VN050067.D
 Acq On : 26 Jul 2018 1:33
 Operator : MD\SY
 Sample : J3821-10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 HR-06-072418-B

Quant Time: Jul 26 07:53:49 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 24 06:27:04 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	613701	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1029019	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	855047	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	299447	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	438252	51.18	ug/l	0.00
Spiked Amount			Recovery	=	102.36%	
35) Dibromofluoromethane	7.60	113	391154	46.56	ug/l	0.00
Spiked Amount			Recovery	=	93.12%	
50) Toluene-d8	10.09	98	1393645	45.12	ug/l	0.00
Spiked Amount			Recovery	=	90.24%	
62) 4-Bromofluorobenzene	12.40	95	375172	36.89	ug/l	0.00
Spiked Amount			Recovery	=	73.78%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	28200	9.81	ug/l	97
18) Methyl Acetate	4.34	43	28864	2.12	ug/l	99
20) Methylene Chloride	4.55	84	70697	8.27	ug/l	98
25) 2-Butanone	6.85	43	8085	2.48	ug/l	90
29) Tetrahydrofuran	7.22	42	2651	1.26	ug/l #	77
43) Isopropyl Acetate	8.17	43	305076	19.33	ug/l #	90
59) 2-Hexanone	10.77	43	16950	8.02	ug/l #	57
95) Naphthalene	15.14	128	77781	9.60	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN072518\
Data File : VN050067.D
Acq On : 26 Jul 2018 1:33
Operator : MD\SY
Sample : J3821-10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
HR-06-072418-B

Quant Time: Jul 26 07:53:49 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 24 06:27:04 2018
Response via : Initial Calibration

