

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072518\
 Data File : VN050071.D
 Acq On : 26 Jul 2018 3:12
 Operator : MD\SY
 Sample : J4124-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-11-D

Manual Integrations
 APPROVED

MMDadoda
 7/26/2018 3:55:15 PM

Quant Time: Jul 26 08:02:35 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 24 06:27:04 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	624416	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1034123	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	856574	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	346249	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	447033	51.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.62%	
35) Dibromofluoromethane	7.59	113	395251	46.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.62%	
50) Toluene-d8	10.09	98	1378101	44.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.80%	
62) 4-Bromofluorobenzene	12.40	95	397358	38.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.76%	
Target Compounds						
16) Acetone	3.83	43	19066	5.04	ug/l	95
18) Methyl Acetate	4.34	43	52532	5.12	ug/l	99
20) Methylene Chloride	4.55	84	446036	57.94	ug/l	99
25) 2-Butanone	6.86	43	14086m	4.24	ug/l	
59) 2-Hexanone	10.76	43	4665	5.95	ug/l	77
95) Naphthalene	15.14	128	405311	27.78	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072518\
Data File : VN050071.D
Acq On : 26 Jul 2018 3:12
Operator : MD\SY
Sample : J4124-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-11-D

Manual Integrations
APPROVED

MMDadoda
7/26/2018 3:55:15 PM

Quant Time: Jul 26 08:02:35 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 24 06:27:04 2018
Response via : Initial Calibration

