


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN072518\
Data File : VN050035.D
Acq On : 25 Jul 2018 7:56
Operator : MD\SY
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
Title : SW846 8260
Last Update : Thu Jul 26 18:10:10 2018



AutoFind: Scans 3362, 3363, 3364; Background Corrected with Scan 3348

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.2	9127	PASS
75	95	30	60	49.6	24920	PASS
95	95	100	100	100.0	50258	PASS
96	95	5	9	6.5	3258	PASS
173	174	0.00	2	1.7	795	PASS
174	95	50	100	93.7	47088	PASS
175	174	5	9	7.5	3525	PASS
176	174	95	101	98.7	46464	PASS
177	176	5	9	7.0	3274	PASS