

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072519\
 Data File : VN056905.D
 Acq On : 24 Jul 2019 11:48
 Operator : JC/SP
 Sample : VSSTDICC002
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSSTDICC002

Quant Time: Jul 25 01:13:34 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072519SIM-VINYL.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 25 01:08:06 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	51052	5.00	ug/l	0.00
4) 1,4-Difluorobenzene	8.58	114	84472	5.00	ug/l	0.00
8) Chlorobenzene-d5	11.41	117	49044	5.00	ug/l	0.00
9) 1,4-Dichlorobenzene-d4	13.35	152	7427	5.00	ug/l	0.00

System Monitoring Compounds

3) 1,2-Dichloroethane-d4	8.01	65	11126	1.83	ug/l	0.00
Spiked Amount	5.000	Range	81 - 118	Recovery	=	36.60%#
5) Dibromofluoromethane	7.59	113	9251	2.03	ug/l	0.00
Spiked Amount	5.000	Range	80 - 119	Recovery	=	40.60%#
6) Toluene-d8	10.09	98	18692	1.75	ug/l	0.00
Spiked Amount	5.000	Range	89 - 112	Recovery	=	35.00%#
7) 4-Bromofluorobenzene	12.40	95	4677	1.28	ug/l	0.00
Spiked Amount	5.000	Range	85 - 114	Recovery	=	25.60%#

Target Compounds

	R.T.	QIon	Response	Conc	Units	Qvalue
2) Vinyl Chloride	2.17	62	8694	1.810	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072519\
Data File : VN056905.D
Acq On : 24 Jul 2019 11:48
Operator : JC/SP
Sample : VSSTDICC002
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSSTDICC002

Quant Time: Jul 25 01:13:34 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072519SIM-VINYL.M
Quant Title : SW846 8260
QLast Update : Thu Jul 25 01:08:06 2019
Response via : Initial Calibration

