

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN072519\
 Data File : VN056913.D
 Acq On : 24 Jul 2019 15:26
 Operator : JC/SP
 Sample : K3921-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SU19-MSW0619-9002

Quant Time: Jul 25 01:33:32 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072519SIM-VINYL.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 25 01:27:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	49216	5.00	ug/l	0.00
4) 1,4-Difluorobenzene	8.58	114	79892	5.00	ug/l	0.00
8) Chlorobenzene-d5	11.41	117	48787	5.00	ug/l	0.00
9) 1,4-Dichlorobenzene-d4	13.35	152	7683	5.00	ug/l	0.00

System Monitoring Compounds						
3) 1,2-Dichloroethane-d4	8.01	65	27560	4.85	ug/l	0.00
Spiked Amount	5.000	Range	81 - 118	Recovery	=	97.00%
5) Dibromofluoromethane	7.59	113	22628	4.90	ug/l	0.00
Spiked Amount	5.000	Range	80 - 119	Recovery	=	98.00%
6) Toluene-d8	10.09	98	46672	4.93	ug/l	0.00
Spiked Amount	5.000	Range	89 - 112	Recovery	=	98.60%
7) 4-Bromofluorobenzene	12.40	95	12233	5.00	ug/l	0.00
Spiked Amount	5.000	Range	85 - 114	Recovery	=	100.00%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN072519\
Data File : VN056913.D
Acq On : 24 Jul 2019 15:26
Operator : JC/SP
Sample : K3921-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SU19-MSW0619-9002

Quant Time: Jul 25 01:33:32 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072519SIM-VINYL.M
Quant Title : SW846 8260
QLast Update : Thu Jul 25 01:27:20 2019
Response via : Initial Calibration

